



## Resetting 400MHz NMR Spectrometer

### Background

This technique will allow you to correct some of the standard issues arising from when a prior user has left the spectrometer unusable for your sample.

### Procedure

1. Insert your sample in the spectrometer.
2. Select the correct deuterated solvent on the Standard tab.
3. In the command line type **ResetInst** and press enter.
4. On the Lock tab press **Find Z0** and let it complete.
5. On the Lock tab press the **Gradient Shim** and let it complete.

**You should see that the instrument shims to <1% rms error and there is a good lock signal.**

The **ResetInst** macro will load a shim set that is stored by management and updated when needed. It will also reset standard instrument parameters back to facility standards.